

MAR 15 2011

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130185</u> ✓	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>TODD BRUNO</u>	ID# <u>7672</u> ✓ Date <u>3/8/11</u> ✓
A Agency <u>NPS GUAKER BAY NP</u> Phone # <u>697 2687</u>	
Instrument Location <u>BARTLET LOVE RANGER STATION</u>	
B Alco S/N <u>2K</u> ✓ Target Value <u>.080</u> ✓ High Pressure <u>800 psi</u>	
Alco Test Values <u>0.079</u> ✓ <u>0.079</u> ✓	
1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of APRIL, 2011 AN

[Signature]
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

130185 ^{8mb} 4/1/11

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

PRINTER TEST